Approved For Refer to 2001/08/05 - CHARDP81-01044R000100010016-4

CENTRAL INTELLIGENCE AGENCY

REPORT NO

CD NO.

(6

INFORMATION REPORT CO

COUNTRY Hungary

SUBJECT Examination of Glass Pharmaceutical Vial

DATE DISTR. 6

January 51

NO. OF PAGES

ī

PLACE ACQUIRED DATE ACQUIRED DATE OF TO CIA LIBRARY NO. OF ENCLS.

SUPPLEMENT TO REPORT NO.

25X1A

25X1X

one from a consignment snipped by Elektroimpex, Budapest, Budapest

- 1. The vial was subjected to visual examination and spectro-analysis.
- 2. Evaluation of the vial is as follows:
 - a. The quality and workmanship are comparable with similar USA products.
 - b. The spectro-analysis reveals a composition acceptable for glass ampoules by USA standards.
 - c. The properties of the glass closely match those of a South American, amber, ampoule glass tested by Corning several years ago.
 - d. The ampoule was fabricated from machine-made glasshing tubing.
 - e. The evaluating laboratory states that there is nothing remarkable about the ampoule or the glass from which it is made and that they are acceptable by US standards.

- end -

CLASSIFICATION STORE

STATE NAVY NSRB DISTRIBUTION
ARMY AIR FBI HFC